QUALITY CONTROL CERTIFICATE

BACITRACIN LOW Diagnostic Tablets

REF:	40211
LOT:	210058
Date of analysis:	2022-04
Date of expiry:	2025-04
Appearance:	Plane tablet, 6 mm in diameter. Complies.
Average weight:	110 mg. Complies.

Performance:

Test strains:

Streptococcus bovis ATCC 15351
Streptococcus agalactiae ATCC 12386
Streptococcus pyogenes ATCC 12344
Streptococcus agalactiae nr. 623

Product complies with specification.

Rosco Diagnostica A/S

Peter Samuelsen
Manager Quality Assurance

ROSCO DIAGNOSTICA A/S · TAASTRUPGAARDSVEJ 30 · DK-2630 TAASTRUP · DENMARK
Phone: +45 43 99 33 77 · Telefax: +45 43 52 73 74 · e-mail: info@rosco.dk