

QUALITY CONTROL CERTIFICATE

CLOXACILLIN 500 µg Diagnostic Tablets

REF: 10311

LOT: 210066

Date of analysis: 2022-04

Date of expiry: 2025-04

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 25922

E. coli 4396

E. Cloacae 4916

Test tablets:

Ceftazidime

Cefotaxitin

Product complies with specification.

Rosco Diagnostica a/s



Peter Samuelsen

Quality Assurance Manager