

QUALITY CONTROL CERTIFICATE

PHENYLBORONIC ACID Diagnostic Tablets

REF: 10411

LOT: 230024

Date of analysis: 2024-02

Date of expiry: 2027-12

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Klebsiella pneumoniae KPC
E. coli NEQAS 4396
E. cloacae FN 9475
E. coli ATCC 25922

Product complies with specification.

A/S Rosco



Niels Frimodt-Møller
Manager Quality Assurance

Rosco Diagnostica A/S