

QUALITY CONTROL CERTIFICATE

BACITRACIN LOW Diagnostic Tablets

REF: 40211
LOT: 230037
Date of analysis: 2023-07
Date of expiry: 2026-07
Appearance: Plane tablet, 6 mm in diameter.
Complies.
Average weight: 110 mg.
Complies.

Performance:

Test strains:

Streptococcus bovis ATCC 15351
Streptococcus agalactiae ATCC 12386
Streptococcus pyogenes ATCC 12344
Streptococcus agalactiae nr. 623

Product complies with specification.

Rosco Diagnostica A/S



Niels Frimodt-Møller
Manager Quality Assurance