

QUALITY CONTROL CERTIFICATE

**AMOXYCILLIN+CLAVULANATE 20+10 µg Neo-Sensitabs**

REF: 60112

LOT: 230015

Date of analysis: 2023-10

Date of expiry: 2025-10

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*E. coli* ATCC 25922

*E. coli* ATCC 35218

Haem.influ. ATCC 49247

*Staphylococcus aureus* ATCC 25923

*Enterococcus faecalis* ATCC 29212

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

A/S Rosco



Niels Frimodt-Møller  
Manager Quality Assurance