

## QUALITY CONTROL CERTIFICATE

### AZITHROMYCIN 15 µg Neo-Sensitabs

REF: 60312

LOT: 220080

Date of analysis: 2023-05

Date of expiry: 2028-04

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*Streptococcus pneumoniae* ATCC 49619  
*Staphylococcus aureus* ATCC 29213

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Niels Frimodt-Møller  
Manager Quality Assurance