

QUALITY CONTROL CERTIFICATE

NALIDIXAN 30 µg Neo-Sensitabs

REF: 61412

LOT: 220022

Date of analysis: 2022-08

Date of expiry: 2026-08

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 25922

Haemophilus influenzae ATCC49766

Klebsiella pneumoniae ATCC13883

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelson
Manager Quality Assurance