

QUALITY CONTROL CERTIFICATE

TEICOPLANIN 30 µg Neo-Sensitabs

REF: 61912

LOT: 220013

Date of analysis: 2022-07

Date of expiry: 2027-06

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Streptococcus pneumonia ATCC 49619

Staphylococcus aureus ATCC 29213

Enterococcus faecalis 29212

Staphylococcus aureus ATCC 700788

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelson
Manager Quality Assurance