

QUALITY CONTROL CERTIFICATE

TRIMETHOPRIM+SULFA 1.25+23.75 µg Neo-Sensitabs (Trimethoprim+Sulfamethoxazole)

REF: 62212

LOT: 220056

Date of analysis: 2023-02

Date of expiry: 2027-03

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 25922
Enterococcus faecalis ATCC 29212
Staphylococcus aureus ATCC 29213
Streptococcus pneumoniae ATCC 49619
Haemophilus influenzae NCTC 8468

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelson
Manager Quality Assurance