

QUALITY CONTROL CERTIFICATE

RIFAMPICIN 5 µg Neo-Sensitabs

REF: 62712

LOT: 210071

Date of analysis: 2022-05

Date of expiry: 2027-05

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Staptococcus pneumonia ATCC 49619

Haemophilus influezae ATCC 49766

Staphylococcus aureus ATCC 29213

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Rosco Diagnostica A/S



Peter Samuelsen
Manager Quality Assurance