

## QUALITY CONTROL CERTIFICATE

### **CEFEPIME 30 µg Neo-Sensitabs**

REF: 63712

LOT: 220037

Date of analysis: 2022-11

Date of expiry: 2025-10

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*E. coli* ATCC 25922

*Pseudomonas aeruginosa* ATCC 27853

*Strep. Pneumonia* ATCC 49619

*Haem. Infl.* ATCC 49766

*Staphylococcus aureus* ATCC 25923

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelsen  
Manager Quality Assurance