

QUALITY CONTROL CERTIFICATE

GENTAMICIN 250 µg Neo-Sensitabs

REF: 63512

LOT: 230013

Date of analysis: 2023-10

Date of expiry: 2028-10

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Enterococcus faecalis ATCC 51299

Enterococcus faecalis ATCC 29212

Enterococcus faecalis ATCC 33186

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

A/S Rosco



Niels Frimodt-Møller
Manager Quality Assurance