

QUALITY CONTROL CERTIFICATE

KANAMYCIN 500 µg Neo-Sensitabs

REF: 63812

LOT: 220078

Date of analysis: 2023-05

Date of expiry: 2028-04

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 25922

Staphylococcus aureus ATCC 25923

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Niels Frimodt-Møller
Manager Quality Assurance