

QUALITY CONTROL CERTIFICATE

CEFTRIAXONE 30 µg Neo-Sensitabs

REF: 64212

LOT: 230027

Date of analysis: 2024-02

Date of expiry: 2028-02

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 25922

Streptococcus pneumoniae ATCC 49619

Haemophilus influenzae ATCC 49766

Ps. Aeruginosa ATCC 27853

S. aureus ATCC 25923

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

A/S Rosco



Niels Frimodt-Møller
Manager Quality Assurance