

## QUALITY CONTROL CERTIFICATE

### **MEROPENEM 10 µg Neo-Sensitabs**

REF: 64312

LOT: 230051

Date of analysis: 2023-12

Date of expiry: 2026-12

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*E. coli* ATCC 25922

*St.pn* ATCC 49619

*H. influn* ATCC49766

*Pseudomonas aeruginosa* ATCC 27853

*S.aur.* ATCC 25923

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

A/S Rosco



Niels Frimodt-Møller  
Manager Quality Assurance