

QUALITY CONTROL CERTIFICATE

CEFTAZIDIME+CLAVULANATE 30+10 µg Neo-Sensitabs

REF: 64612

LOT: 210048

Date of analysis: 2022-01

Date of expiry: 2024-01

Appearance: Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 35218

E.coli ATCC 25922

Klebsiella pneumoniae ATCC 700603

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelsen
Manager Quality Assurance