



QUALITY CONTROL CERTIFICATE

CEFOTAXIME+CLAVULANATE 30+10 µg Neo-Sensitabs

REF: 64712

LOT: 220083

Date of analysis: 2023-04

Date of expiry: 2025-04

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 35218

E. coli ATCC 25922

Klebsiella pneumoniae ATCC 700603

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Rosco Diagnostica A/S



Niels Frimodt-Møller
Manager Quality Assurance