

QUALITY CONTROL CERTIFICATE

CEFEPIME+CLAVULANATE 30+10 µg Neo-Sensitabs

REF: 64812

LOT: 210062

Date of analysis: 2022-03

Date of expiry: 2024-03

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 35218

E. coli ATCC 25922

Klebsiella pneumoniae ATCC 700603

Ent.aerogens NEQAS 1057

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelson
Manager Quality Assurance