

QUALITY CONTROL CERTIFICATE

AMPICILLIN 2 µg Neo-Sensitabs

REF: 66712

LOT: 220061

Date of analysis: 2023-03

Date of expiry: 2028-03

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Ent. Faecalis ATCC29212
Strep. Pneumonia ATCC49619
Haemophilus influenza ATCC49766
Staph.Aureus ATCC29213

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelsen
Manager Quality Assurance