QUALITY CONTROL CERTIFICATE

NITROFURANTOIN 100 µg Neo-Sensitabs

REF:	67012
LOT:	220053
Date of analysis:	2023-02
Date of expiry:	2028-02
Appearance:	Light yellow, plane tablet, 9 mm in diameter, on both sides printed with a black identification code. Complies.
Thickness:	Mean of 10: 1.50-1.65 mm. Complies.

Performance:

Test strains:

Streptococcus pneumoniae ATCC 49619
E. coli ATCC 25922
Enterococcus faecalis ATCC 29212
Staphylococcus aureus ATCC 29213

Product complies with specifications.

Uniformity:	Inhibition zones of individual tablets vary less than ± 1 mm from mean. Complies.
Microbiological quality:	Complies.

Manager Quality Assurance

Rosco Diagnostica A/S

Peter Samuelsen
Manager Quality Assurance

ROSCO DIAGNOSTICA A/S · TAASTRUPGAARDSVEJ 30 · DK-2630 TAASTRUP · DENMARK
Phone: +45 43 99 33 77 · Telefax: +45 43 52 73 74 · e-mail: info@rosco.dk