

## QUALITY CONTROL CERTIFICATE

### AMOXYCILLIN+CLAVULANATE 2+1 µg Neo-Sensitabs

REF: 69812

LOT: 210078

Date of analysis: 2022-05

Date of expiry: 2024-05

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*E.coli* ATCC 35218

*Haemophilus influenza* ATCC49766

*Staphylococcus aureus* ATCC 29213

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelson  
Manager Quality Assurance