

## QUALITY CONTROL CERTIFICATE

### AMOXYCILLIN 30 µg Neo-Sensitabs

REF: 70112

LOT: 230022

Date of analysis: 2023-09

Date of expiry: 2027-07

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

#### Performance:

Test strains:

*E. coli* ATCC 25922  
*Enterococcus faecalis* ATCC 29212

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Niels Frimodt-Møller  
Manager Quality Assurance