

## QUALITY CONTROL CERTIFICATE

### **BACITRACIN 40 U Neo-Sensitabs**

REF: 70812

LOT: 220045

Date of analysis: 2022-11

Date of expiry: 2027-11

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*Enterococcus faecalis* ATCC 29212  
*Staphylococcus aureus* ATCC 25923

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelsen  
Manager Quality Assurance