

## QUALITY CONTROL CERTIFICATE

### LEVOFLOXACIN 5 µg Neo-Sensitabs

REF: 74812

LOT: 220019

Date of analysis: 2022-09

Date of expiry: 2026-08

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains: *E. coli* ATCC 25922  
*Staphylococcus aureus* ATCC 29213  
*Pseudomonas aeruginosa* ATCC 27853  
*Staphylococcus aureus* ATCC 25923  
*Streptococcus pneumoniae* ATCC 49619  
*Ham.influ* ATCC 49766

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelsen  
Manager Quality Assurance