

QUALITY CONTROL CERTIFICATE

NORFLOXACIN 10 µg Neo-Sensitabs

REF: 76212

LOT: 220073

Date of analysis: 2023-04

Date of expiry: 2027-04

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains: *E. coli* ATCC 25922
Staphylococcus aureus ATCC 25923
Pseudomonas aeruginosa ATCC 27853
Ent.fac. ATCC29212
Staphylococcus aureus ATCC 29213

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Niels Frimodt-Møller
Manager Quality Assurance