

QUALITY CONTROL CERTIFICATE

MOXIFLOXACIN 5 µg Neo-Sensitabs

REF: 79812

LOT: 220079

Date of analysis: 2023-05

Date of expiry: 2028-04

Appearance: Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 25922
Staphylococcus aureus ATCC 29213
Strep. pneum. ATCC 49619
Hem. Infl. ATCC 49766


Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Niels Frimodt-Møller
Manager Quality Assurance