

QUALITY CONTROL CERTIFICATE

TEMOCILLIN 30 µg Neo-Sensitabs

REF: 78312

LOT: 230068

Date of analysis: 2024-02

Date of expiry: 2027-01

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 25922

E. coli ATCC 35218

K. pneumonia 7759

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

A/S Rosco



Niels Frimodt-Møller
Manager Quality Assurance