

## QUALITY CONTROL CERTIFICATE

### **ERTAPENEM 10 µg Neo-Sensitabs**

REF: 80712

LOT: 210045

Date of analysis: 2022-01

Date of expiry: 2025-12

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*E. coli* ATCC 25922

*Streptococcus pneumonia* ATCC 49619

*Hemophilus influenza* ATCC 49766

*Staphylococcus aureus* ATCC 29213

*Pseudomonas aeruginosa* ATCC 27853

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelson  
Manager Quality Assurance