

## QUALITY CONTROL CERTIFICATE

### **AMPHOTERICIN B 10 µg Neo-Sensitabs**

REF: 81012

LOT: 230020

Date of analysis: 2023-12

Date of expiry: 2026-12

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*Candida albicans* ATCC 90028  
*Candida parapsolis* ATCC 22019  
*Candida krusei* ATCC 6258

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Manager Quality Assurance

Rosco Diagnostica A/S



Niels Frimodt-Møller  
Manager Quality Assurance