

QUALITY CONTROL CERTIFICATE

AmpC Confirm ID Pack, Neo-Sensitabs

REF: 98007

LOT: 220029

Date of expiry: 2024-04

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Test strains	Results
<i>E. coli</i> ATCC 25922	Negative
<i>Ent. Cloacae</i> NCTC 13406	Amp C Positive
<i>Ent. Cloacae</i> ATCC BAA - 1143	Amp C Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelson
Manager Quality Assurance