

QUALITY CONTROL CERTIFICATE

Urin Kit Privat Praksis Neo-Sensitabs + Diatab

REF: 98009

LOT: 1903-2

Date of expiry: 2023-05

Appearance: NeoSensitab is Light yellow, plane tablet, 9 mm in diameter, on both sides printed with a black identification code. Diatab is a plane tablet, 6 mm in diameter. Complies.

Thickness: Mean of 10: 1.50-1.65 mm. (Neosensitabs)
Average weight: 110 mg (Diatabs)
Complies.

Performance:

Test strains:

<i>Staphylococcus aureus</i>	ATCC 29213
<i>Klebsiella pneumoniae</i>	7323
<i>H. influenza</i>	NCTC 8468
<i>E. coli</i>	ATCC 25922
<i>Enterococcus faecalis</i>	ATCC 29212
<i>Staphylococcus aureus</i>	ATCC 25923
<i>Pseudomonas aeruginosa</i>	ATCC 27853
<i>Klebsiella pneumoniae</i>	ATCC 13883
<i>Staphylococcus simulans</i>	ATCC 27848
<i>Staphylococcus haemolyticus</i>	88140
<i>S. haemolyticus</i>	ATCC 29970

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than ± 1 mm from mean. Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelson
Manager Quality Assurance