

QUALITY CONTROL CERTIFICATE

ESBL Confirmation (Chrom.AmpC) kit

REF: 98012

LOT: 230001

Date of expiry: 2024-03

Appearance: Light yellow, plane tablets, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains	Results
<i>E. coli</i> ATCC 25922	Negative
<i>K. pneumone</i> ATCC 700603	ESBL Positive
<i>Ent. cloacae</i> NCTC 13464	ESBL Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Peter Samuelson
Manager Quality Assurance