

## QUALITY CONTROL CERTIFICATE

### **Total ESBL Confirmation Kit**

REF: 98014

LOT: 230091

Date of expiry: 2024-10

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains	Results
<i>E. coli</i> ATCC 25922	Negative
<i>K. pneumonia</i> ATCC 700603	ESBL Positive
ENT. Cloacae NCTC 13464	ESBL Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Niels Frimodt-Møller  
Manager Quality Assurance