

QUALITY CONTROL CERTIFICATE

KPC, MBL and OXA-48 Kit, Neo-Sensitabs

REF: 98015

LOT: 240001

Date of expiry: 2025-07

Appearance: Light yellow, plane tablets, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains	Results
<i>E. coli</i> ATCC 25922	Negative
<i>K. pneumoniae</i> NCTC 13438	KPC Positive
<i>K. pneumoniae</i> ATCC BAA – 1705	KPC Positive
<i>K. pneumoniae</i> ATCC BAA – 2146	MBL Positive
<i>K. pneumoniae</i> NCTC 13439	MBL Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

A/S Rosco



Niels Frimodt-Møller
Manager Quality Assurance