

## QUALITY CONTROL CERTIFICATE

### **Total ESBL, AmpC and ESBL+AMPC Confirm Kit, Neosensitabs**

REF: 98019

LOT: 230089

Date of expiry: 2024-06

Appearance: Light yellow, plane tablets, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains

<i>E. coli</i> ATCC 25922	Negative
<i>K. pneumoniae</i> ATCC 700603	ESBL Positive
<i>Ent. cloaceae</i> NCTC 13406	AmpC Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Niels Frimodt-Møller  
Manager Quality Assurance