

QUALITY CONTROL CERTIFICATE

NEO-Rapid CARB Kit Diagnostic Tablets

REF: 98024

LOT: 23004

Date of expiry: 2024-05

Appearance: Plane tablet, 6 mm in diameter.
Complies.

Average weight: 90 mg.
Complies.

Test strains

DIATABS	Positive	Negative
Imipenem(x2) + Indicator (CARB)	<i>Klebsiella pneumoniae</i> BAA 1705	<i>E. coli</i> ATCC 25922

Product complies with specifications.

Rosco Diagnostica A/S



Peter Samuelson
Manager Quality Assurance