

QUALITY CONTROL CERTIFICATE

NITRATE REDUCTION Diagnostic Tablets

REF: 43711
LOT: 1601-1
Date of analysis: 2016-08
Date of expiry: 2020-07
Appearance: Plane tablet, 6 mm in diameter.
Complies.
Average weight: 90 mg.
Complies.

Performance:

Test strains:

E. coli ATCC 25922
Staphylococcus saprophyticus ATCC 15305
Staphylococcus aureus ATCC 25923
Staphylococcus saprophyticus 27842 FN 3784
Acinetobacter. Calcoaceticus ATCC 23055

Product complies with specification.

Rosco Diagnostica A/S



Michela Venturi
Manager Quality Assurance