

## QUALITY CONTROL CERTIFICATE

### **Piperacillin 100 µg Neo-Sensitabs**

REF: 77312

LOT: 1601-2

Date of analysis: 2016-03

Date of expiry: 2020-03

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

### Performance:

Test strains:

*E. coli* ATCC 25922

*Enterococcus faecalis* ATCC 29212

*Pseudomonas aeruginosa* ATCC 27853

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica A/S



Michela Venturi  
Manager Quality Assurance