

QUALITY CONTROL CERTIFICATE

CEFAZOLIN 30 µg Neo-Sensitabs

REF: 60412
LOT: 230044
Date of analysis: 2024-06
Date of expiry: 2028-05
Appearance: Complies
Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 25922
Staphylococcus aureus ATCC 25923
E.Coli ATCC 35218

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality