

## QUALITY CONTROL CERTIFICATE

### **PENICILLIN 10 U Neo-Sensitabs**

REF: 61712

LOT: 230038

Date of analysis: 2024-05

Date of expiry: 2028-03

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

#### Performance:

Test strains:  
*Staphylococcus aureus* ATCC 29213  
*Streptococcus pneumoniae* ATCC 49619

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA  
Chief Medical Officer, Head of Quality