

QUALITY CONTROL CERTIFICATE

NITROFURANTOIN 100 µg Neo-Sensitabs

REF: 67012

LOT: 230042

Date of analysis: 2024-04

Date of expiry: 2029-04

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Streptococcus pneumoniae ATCC 49619

E. coli ATCC 25922

Enterococcus faecalis ATCC 29212

Staphylococcus aureus ATCC 29213

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Manager Quality Assurance

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality
Manager Quality Assurance