

QUALITY CONTROL CERTIFICATE

CEFOTAXIME 5 µg Neo-Sensitabs

REF: 65812

LOT: 230077

Date of analysis: 2024-05

Date of expiry: 2028-03

Appearance: Light yellow, plane tablet, 9 mm in diameter, on both sides printed with a black identification code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

E. coli ATCC 25922

Streptococcus pneumoniae ATCC 45615

Kleb.pn. ATCC700603

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS

A handwritten signature in blue ink that reads "Claus Bo Svendsen".

Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality