# QUALITY CONTROL CERTIFICATE

**CEFOXITIN 30 µg Neo-Sensitabs**

| | |
|---|---|
| REF: | 62912 |
| LOT: | 230076 |
| Date of analysis: | 2024-06 |
| Date of expiry: | 2028-06 |
| Appearance: | Light yellow, plane tablet, 9 mm in diameter, on both sides printed with a black identification code.<br>Complies. |
| Thickness: | Mean of 10: 1.50-1.65 mm.<br>Complies. |

Performance:

Test strains:

*E. coli* ATCC 25922
*Staphylococcus aureus* ATCC 43300
*Staphylococcus aureus* ATCC 29213

Product complies with specifications.

| | |
|---|---|
| Uniformity: | Inhibition zones of individual tablets vary less than ± 1 mm from mean.<br>Complies. |

.

| | |
|---|---|
| Microbiological quality: | Complies. |

Rosco Diagnostica ApS

Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality

**ROSCO DIAGNOSTICA ApS** · STENSMOSEVEJ 24A · DK-2620 ALBERTSLUND · DENMARK
Phone: +45 93 40 65 65 · e-mail: info@rosco-diagnostica.com