

QUALITY CONTROL CERTIFICATE

NOVOBIOCIN 5 µg Neo-Sensitabs

REF: 76312

LOT: 230030

Date of analysis: 2024-05

Date of expiry: 2028-02

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Staphylococcus aureus ATCC 25923
Staphylococcus aureus ATCC 29213

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality