

QUALITY CONTROL CERTIFICATE

Total ESBL, AmpC and ESBL+AMPC Confirm Kit, Neosensitabs

REF: 98019

LOT: 24-4003

Date of expiry: 2025-06

Appearance: Light yellow, plane tablets, 9 mm in diameter, on both sides printed with a black identification code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains

<i>E. coli</i> ATCC 25922	Negative
<i>K. pneumoniae</i> ATCC 700603	ESBL Positive
<i>Ent. cloacae</i> NCTC 13406	AmpC Positive

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality