

-1

## QUALITY CONTROL CERTIFICATE

### **Colistin 10 µg Neo-Sensitabs**

REF: 66312

LOT: 230045-1

Date of analysis: 2024-06

Date of expiry: 2029-05

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

Performance:

Test strains:

*E. coli* ATCC 25922

*Pseudomonas aeruginosa* ATCC 27853

*S. aureus* ATCC 25923

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS

A handwritten signature in blue ink that reads "Claus Bo Svendsen".

Claus Bo Svendsen MD PhD MBA  
Chief Medical Officer, Head of Quality