

QUALITY CONTROL CERTIFICATE

Total ESBL Confirmation Kit

REF: 98014

LOT: 24-4005

Date of expiry: 2025-08

Appearance: Light yellow, plane tablets, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

| Test strains | Results |
|---------------------------------|---------------|
| <i>E. coli</i> ATCC 25922 | Negative |
| <i>K. pneumonia</i> ATCC 700603 | ESBL Positive |
| ENT. Cloacae NCTC 13464 | ESBL Positive |

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality