

QUALITY CONTROL CERTIFICATE

**TRIMETHOPRIM+SULFA 1.25+23.75 µg Neo-Sensitabs
(Trimethoprim+Sulfamethoxazole)**

REF: 62212

LOT: 24-1003

Date of analysis: 2024-07

Date of expiry: 2028-07

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

- E. coli* ATCC 25922
- Enterococcus faecalis* ATCC 29212
- Staphylococcus aureus* ATCC 29213
- Streptococcus pneumoniae* ATCC 49619
- Haemophilus influenzae* NCTC 8468

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality