

QUALITY CONTROL CERTIFICATE

FOSFOMYCIN 200 µg Neo-Sensitabs

REF: 62312

LOT: 24-1009

Date of analysis: 2024-07

Date of expiry: 2027-07

Appearance: Light yellow, plane tablet, 9 mm in diameter, on both sides printed with a black identification code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:
E. coli ATCC 25922
Staphylococcus aureus ATCC 29213
Pseudomonas aeruginosa ATCC 27853

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA
Chief Medical Officer, Head of Quality