

QUALITY CONTROL CERTIFICATE

Rapid ESBL Screen Kit Diagnostic Tablets

REF: 98022
 LOT: 24-4007
 Date of expiry: 2027-07
Appearance: Plane tablet, 6 mm in diameter.
 Complies.

Average weight: 90 mg.
 Complies.

Test strains

DIATABS	Positive	Negative
	<i>Enterobacter cloacae</i> NCTC 13464	<i>E. coli</i> ATCC 25922
Cefotaxime + Indicator	Yellow	Red
Cefotaxime + Tazo + indicator	Red	Red

Product complies with specificat

Rosco Diagnostica ApS



Claus Bo Svendsen MD PhD MBA
 Chief Medical Officer, Head of Quality